

Transport Tubes Instructions For Use (IFU)

Part Numbers: 230417, 230418, 230419, 230419R

To ensure that CELLTREAT brand transport tubes perform to IATA leakproof and shatterproof compliance standards (95 KPa Packaging Instruction 650), confirm that the threading of the caps and tubes are aligned, and that the cap is secured tightly to the tube. Hand-tightness is sufficient to achieve full closure.